



Department: Calibration Laboratory ISO/IEC 17025:2005 Accreditation No. 209

Name and Address:

**Laboratory name** Madey Vered Laboratory  
**Address** 8 Ashlegan Street, Kiryat Gat, 82021, Israel  
**Phone** +972-8-6818899  
**Fax** +972-8-6818801  
**E-Mail** [info@mv-technology.com](mailto:info@mv-technology.com)



Department Calibration Laboratory ISO/IEC 17025:2005 Accreditation No. 209

### Schedule of Accreditation

Item	Scope Type	Site	Measurand, Instrument, Gauge	Range [Including margins] (Does not include margins)	CMC Expressed as an Expanded Uncertainty (95%)	Reference Document	Remarks
<b>Calibration – Big Volume Volumetric Instruments</b>					<b>כיוול - כיוול מכשירים ולומקטריים – נפחים גדולים</b>		
1	A	P	Volume passed, Water	נפח זרימה, מים, Volume: [5 L to 100 L] Flow rate: [25 L/h to 120 L/h]	0.80 %	כללי מדידת מים (מדוי-מים), התשמ"ח – 1988 ת"י 63 חלק 3	Test Bench No.10 Synchronized mass method
2	A	P	Water meters Diameter DN15 to DN25	מדי מים קוטר DN15 to DN25 Volume: [15 L to 100 L] Flow rate: [120 L/h to 1500 L/h]	0.40 %	הוראות עבודה QW-16	May be conducted on test benches No.1, as well by mass method, providing higher uncertainties.
3	A	P		Volume: 100 L Flow rate: (1500 L/h to 7000 L/h]	0.65 %		
4	A	P	Volume passed, Water	נפח זרימה, מים, Volume: [20 L to 100 L] Flow rate: [70 L/h to 280 L/h]	2.2 %	כללי מדידת מים (מדוי-מים), התשמ"ח – 1988 ת"י 63 חלק 3	Test Bench No.3 Mass method
5	A	P	Water meters Diameter DN20 to DN50	מדי מים קוטר DN20 to DN50 Volume: [100 L to 500 L] Flow rate: [280 L/h to 1500 L/h]	1.5 %	הוראות עבודה QW-15	
6	A	P		Volume: [500 L to 1000 L] Flow rate: (1500 L/h to 31,000 L/h]	2.0 %		
7	A	P	Volume passed Water Water meters Diameter DN25 to DN200	ספיקת נפחית מים מדי מים קוטר DN25 to DN200 Flowrate [0.01 m <sup>3</sup> /h – 250 m <sup>3</sup> /h]	0.4 %	כללי מדידת מים (מדוי-מים), התשמ"ח – 1988 ת"י 63 חלק 3 הוראות עבודה QW-17	Test Bench no. 11 Synchronized mass method (reference weighing system)

Site: P or T or M, P-Permanent, T-Temporary, M-Mobile

Type of Scopes: A- Fixed, C- flexible

Flexible scope in analytical tests: Type of matrix, analytes, experimental systems and/or analytical characteristics may be subject to changes, in accordance with the laboratory's approved and documented procedures. For details, please refer to the list of Accredited Tests, available from the laboratory upon request.

Date of issue 10.07.2018 Version No. 4

Page No. 2 of: 2

The certificate attached is an integral part of the schedule and is numbered identically התעודת המצורפת לנספח זה מהווה חלק בלתי נפרד ממנו ומספרה זהה למספר הנספח